

NOV 14 2003

K033468  
p. 1/2

## **SUMMARY OF SAFETY AND EFFECTIVENESS**

### **COMPANY AND CONTACT PERSON**

Medtronic Perfusion Systems  
7611 Northland Drive  
Minneapolis, MN 55428  
Tel: (763) 391-9183  
FAX: (763) 391-9603

Michele Pyfferoen, Regulatory Affairs Specialist, Regulatory Affairs

### **DEVICE NAME**

AFFINITY® 20 µ Arterial Blood Filter with Trillium™ Biosurface

### **NAME OF PREDICATED OR LEGALLY MARKETING DEVICE**

AFFINITY® 20 µ Arterial Blood Filter (K994208)  
AFFINITY® Hollow Fiber Oxygenator with Trillium™ Biosurface (K973760)

### **DESCRIPTION OF DEVICE**

The AFFINITY® 20 µ Arterial Blood Filter with Trillium™ Biosurface is a single use device designed to filter microemboli from the blood in the extracorporeal circuits during cardiopulmonary bypass surgery.

The AFFINITY<sup>(R)</sup> 20 µ Arterial Blood Filter with Trillium™ Biosurface is coated with a non-leaching biosurface.

### **STATEMENT OF INTENDED USE**

The AFFINITY® 20 µ Arterial Blood Filter with Trillium™ Biosurface is indicated for use in cardiopulmonary bypass procedures up to 6 hours in duration for the removal of particulate and gaseous microemboli.

### **STATEMENT OF INTENDED USE OF PREDICATED/MARKETING DEVICE**

The AFFINITY® 20 µ Arterial Blood Filter is indicated for use in cardiopulmonary bypass procedures up to 6 hours in duration for the removal of particulate and gaseous microemboli.

### **STATEMENT OF TECHNOLOGICAL CHARACTERISTICS COMPARISON**

Information regarding technological characteristics comparison is provided in the following section, "Determination of Substantial Equivalence".

## **DETERMINATION OF SUBSTANTIAL EQUIVALENCE**

This “**SPECIAL 510(k)**” is being submitted for a modification to the AFFINITY® 20 µ Arterial Blood Filter. The modification to the current AFFINITY® 20 µ Arterial Blood Filter is to coat the blood contact surfaces with Trillium™.

The AFFINITY® 20 µ Arterial Blood Filter with Trillium™ Biosurface is being compared to the following Marketed Devices:

- AFFINITY® 20 µ Arterial Blood Filter (K994208)
- AFFINITY® Hollow Fiber Oxygenator with Trillium™ Biosurface (K973760)

The AFFINITY® 20 µ Arterial Blood Filter with Trillium™ Biosurface Trillium™ has the same indications statement and intended uses as the:

- AFFINITY® 20 µ Arterial Blood Filter (K994208)

The AFFINITY® 20 µ Arterial Blood Filter with Trillium™ Biosurface has “no new technological characteristics (e.g., materials and manufacturing processes)” from the AFFINITY® 20 µ Arterial Blood Filter. The technological characteristic is solely the coating material of the blood contact surface:

- Trillium™

The technological characteristic of the Trillium™ Biosurface is common to other medical devices (hollow fiber oxygenators) currently in commercial distribution as follows:

- AFFINITY® Hollow Fiber Oxygenator with Trillium™ Biosurface (K973760)

This technological characteristic “could affect the safety and effectiveness of the device”. However, these “technological characteristics do not raise new types of safety or effectiveness questions”. In addition, “there are acceptable scientific methods which exist for assessing effects of these new technological characteristics”.

“Performance data to assess the effects of these new technological characteristics” has been performed. These “performance data demonstrate” that the AFFINITY® 20 µ Arterial Blood Filter with Trillium™ Biosurface is substantially equivalent to other marketed arterial filters.

The biocompatibility and *in vitro* bench testing demonstrated that when compared to the predicate devices, the AFFINITY® 20 µ Arterial Blood Filter with Trillium™ Biosurface does not significantly affect safety and effectiveness and are substantially equivalent to other commercially distributed arterial filters. The *in vitro* bench testing included analysis of:

- Coating Characteristics
- Physical Characteristics
- Performance Characteristics



DEPARTMENT OF HEALTH & HUMAN SERVICES

Public Health Service

NOV 14 2003

Food and Drug Administration  
9200 Corporate Boulevard  
Rockville MD 20850

Medtronic Perfusion Systems  
c/o Ms. Michele Pyfferoen  
7611 Northland Drive  
Minneapolis, MN 55428

Re: K033468  
AFFINITY® 20µ Arterial Filter with Trillium™ Biosurface  
Regulation Number: 21 CFR 870.4260  
Regulation Name: Cardiopulmonary Bypass Arterial Line Blood Filter  
Regulatory Class: Class II (two)  
Product Code: 74 DTM  
Dated: October 31, 2003  
Received: November 3, 2003

Dear Ms. Pyfferoen:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to such additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.


Page 2 - Ms. Michele Pyfferoen

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

This letter will allow you to begin marketing your device as described in your Section 510(k) premarket notification. The FDA finding of substantial equivalence of your device to a legally marketed predicate device results in a classification for your device and thus, permits your device to proceed to the market.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801 and additionally 21 CFR Part 809.10 for in vitro diagnostic devices), please contact the Office of Compliance at (301) 594-4646. Additionally, for questions on the promotion and advertising of your device, please contact the Office of Compliance at (301) 594-4639. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). Other general information on your responsibilities under the Act may be obtained from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 443-6597 or at its Internet address <http://www.fda.gov/cdrh/dsma/dsmamain.html>

Sincerely yours,

A handwritten signature in black ink, appearing to read "Bram D. Zuckerman".

Bram D. Zuckerman, M.D.  
Director

Division of Cardiovascular Devices  
Office of Device Evaluation  
Center for Devices and  
Radiological Health

Enclosure

510(k) Number (if known) K033468

Device Name:

**AFFINITY® 20µ Arterial Filter with Trillium™ Biosurface**


Indications for Use:

The AFFINITY® Arterial Filter with Trillium™ Biosurface is indicated for use in cardiopulmonary bypass procedures up to 6 hours in duration for the removal of particulate and gaseous microemboli.

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

---

Concurrence of CDRH, Office of Device Evaluation (ODE)



**(Division Sign-Off)**  
**Division of Cardiovascular Devices**

510(k) Number K033468

(Optional Format 3-10-98)